

PROOF AND TRALLE HYDROMETER

0 to 200 Percent scale is Proof of Spirit Scale.
The 0 to 100 Scale is for percent of alcohol by volume, Tralles.
This Hydrometer is for distilled alcohol only. It cannot be used
for beer, wine, or any other fermented beverage.

TEMPERATURE CORRECTION CHART

Degrees Fahrenheit	Proof 0 to under 25	Proof 25 to under 50	Proof 50 and over	Tralles
100	14	12	16	8
95	12	10.5	14	7
90	10	9	12	6
85	8.5	7.5	10	5
80	7	6	8	4
75	5	4.5	6	3
70	3	3	4	2
65	1.5	1.5	2	1
60	0	0	0	0
55	1.5	1.5	2	1
50	3.5	3	4	2
45	5	4.5	6	3
40	7	6	8	4
35	9	8	10	5
30	10.5	9	12	6
25	12	10.5	14	7
20	14	12	16	8
15	16	13.5	18	9
10	18	15	20	10
5	19	16.5	22	11
0	21	18	24	12

Temperature reading below 60 degrees Fahrenheit add to observed hydrometer reading.

Temperature reading above 60 degrees Fahrenheit subtract from observed hydrometer reading.